

Introducing ProCair by Novis – alternating solutions for the prevention and treatment of pressure injuries and other preventable complications of immobility.

			Standard size	King Single Size
General	CODE		APMPC-R03	APMPC-R03K
	CAPACITY (SAFE WORKING LOAD)		30 - 200 kg	
	SYSTEM ALTERNATION		1in 2	
	WARRANTY		3 year	
	CARRY BAG		Carry bag with shoulder strap	
	COMPLIANCE		IEC 60601-1, IEC60601-1-2, IEC60601-1-11	
	ARTG		289458	
	CARTON DIMENSIONS (L × W × H) in mm		100 x 670 x 270 mm	1150 x 670 x 300 mm
Support Surface	DIMENSIONS (L x W x H) in mm		2000 x 900 x 180 mm	2000 x 1050 x 1800 mm
	MATTRESS WEIGHT		12 kg	13 kg
	WEIGHT SENSOR PAD		No	
	CPR TYPE		QuickTwist CPR at head end	
	CELLS	NUMBER	21	
		STATIC HEAD CELLS / ZONE	3	
		NARROW FOOT CELLS / ZONE	4	
		CELL FOAM INSERT MATERIAL	High resilience PU	
		CELL MATERIAL	TPU laminated nylon	
		MATTRESS AIR SUPPLY LINE	PVC air manifold and CPC quick release connectors	
	TOP COVER	MATERIAL/TYPE	Dark blue, polyurethane-laminated	dnylon
			- water resistant, vapour permeab	le multi-stretch
		FIRE RETARDANCY RATE	BS7175 Crib 5	
		MAXIMUM WASH TEMPERATURE	95° C	
		ATTACHEMENT	Dual zippers	
		SEAM CONSTRUCTION	RF welded on all external seams	
	BASE	MATERIAL / TYPE	Black polyester	
		HINGE	No	
		FIXING STRAPS	Webbing at 4 corners, 2 at middle	





POWER		110 - 220V 50/60Hz			
COMPRESSORS		Single			
OPERATION		Full digital control	Full digital control		
CYCLE TIME		12 minutes	12 minutes		
NOISE LEVEL, WITHIN	I 2 METERS	28db(A) ± 3	28db(A) ± 3		
MIN / MAX PRESSURE		20-60mm HG +/-6mm H	20-60mm HG +/-6mm HG		
ELECTRICAL PROTECT	TION TYPE	Class II Type BF	Class II Type BF		
OPERATING ENVIRONMENT		Operation 15°C to 35°C	Operation 15°C to 35°C Storage 5°C to 60°C		
	AUTOMATIC	Yes – cyclical	Yes – cyclical		
PRESSURE SETTING	MANUAL	5 levels			
CAIRALERT PATIENT EXIT ALARM		No			
FOWLER FUNCTION		No	Yes, Manual		
BACKUP BATTERY		No			
ALADMS		Low Pressure			

Control Unit

ALARMS



Power Failure

THERAPEUTIC DEVICES SHOULD ONLY BE USED IN ACCORDANCE WITH MANUFACTURERS INSTRUCTIONS AND UNDER THE CONSENT, SUPERVISION AND MANAGEMENT OF A SUITABLY QUALIFIED HEALTH PROFESSIONAL.

